

BIOCOMPATIBILITY TESTS PERFORMED

TESTED SAMPLES

RUTHINIUM DISC,

RUTHINIUM DISC RX,

RUTHINIUM DISC

BICOLOUR, RUTHINIUM

DISC NEUTRAL,

RUTHINIUM DISC MULTILAYER,

RUTHINIUM DISC MULTILAYER SC

MANUFACTURER

DENTAL MANUFACTURING SPA RUTHINIUM GROUP VIA CA' MIGNOLA NUOVA, 1699 BADIA POLESINE (RO) - ITALY

It is hereby declared that the medical devices listed above have been pool-tested for biocompatibility according to the international standard ISO 10993-1:2018, in detail:

- Chemical/physical characterisation test according to ISO 10993-18:2020/A1 2022
- Cytotoxicity test according to ISO 10993-5:2009 non-cytotoxic
- In vitro skin irritation test according to ISO 10993-23:2021 non-irritant
- GPMT Magnusson Kligman skin sensitisation test according to ISO 10993-10:2021 non-sensitising.

Issue authorised by: Bologna, 20/03/2023

Laboratory Manager: Dr. Lorenzo

Autore

Lab4LIFE S.r.l.

Lorenzo Santore

info@lab4life.com | lab4lifesrl@legalmail.it | lab4life.com

